

CERTIFICADO DE ANÁLISE

POLISENSIDISC 12 - DME

PSEUDOMONAS/ACINETOBACTER LOTE PPADME10318 FAB. 03/18 VAL. 03/19

Antibacteriano	Potência/ Código	Lote Sensidisc	Pseudomonas aeruginosa ATCC 27853	
			Halo	Halo/esperado (mm)
Amicacina	AMI 30	2632AMI	26,0	18-26
Aztreonam	ATM 30	2628ATM	29,0	23-29
Cefepime	CPM 30	2656CPM	31,0	24-31
Cefotaxima	CTX 30	2579CTX	21,0	18-22
Ceftazidima	CAZ 30	2555CAZ	29,0	22-29
Ciprofloxacina	CIP 05	2624CIP	32,0	25-33
Gentamicina	GEN 10	2620GEN	21,0	17-23
Levofloxacina	LEV 05	2634LEV	24,0	19-26
Norfloxacina	NOR 10	2598NOR	29,0	22-29
Piperacilina/Tazobactam	PIT 110	2609PIT	33,0	25-33
Sulfazotrim	SUT 25	2641SUT	-	-
Tobramicina	TOB 10	2601TOB	25,0	20-26

Informações Adicionais:

- **Meio utilizado:** Mueller Hinton Agar – DIFCO, lote 7193688 com validade em 31/10/2019.

Araçatuba, 03 de Abril de 2018.



Controle de Qualidade – DME
Karina de Campos Salles - CRBM 12001